

**STANDARDS**
MALAYSIA

Certificate of Accreditation

No: SAMM 669

Accredited since: 17 October 2014

This is to certify that

EDMI ELECTRONICS SDN. BHD.
SENAI, JOHOR
MALAYSIAScan QR Code or visit
www.ism.gov.my/cab-directories
for the current scope of accreditation

has been granted accreditation in respect of the scope of accreditation described in the schedule, subject to the terms and conditions governing the *Skim Akreditasi Makmal Malaysia* (SAMM), the Laboratory Accreditation Scheme of Malaysia.

Laboratories accredited under SAMM meet the requirements of MS ISO/IEC 17025. This Malaysian Standard is identical with ISO/IEC 17025 published by the International Organization for Standardization (ISO).



(DATUK FADILAH BAHARIN)
Director General
Department of Standards Malaysia

Date of issue: 6 September 2017

Schedule

Issue date: 2 March 2020
Valid until: 17 October 2023



NO: SAMM 669

Page: 1 of 1

LABORATORY LOCATION:
(PERMANENT LABORATORY)



EDMI ELECTRONICS SDN. BHD.
PLO 226A, JALAN CYBER 1A,
SENAI INDUSTRIAL PARK PHASE III
81400 SENAI, JOHOR
MALAYSIA

FIELD OF TESTING: MECHANICAL

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF TESTING: MECHANICAL

Materials/ Products Tested	Type of Test/ Properties Measured/ Range of Measurement	Standard Test Methods/ Equipment/Techniques
*Electronic Gas Meter	Error of indication for flow rate in the range (0.03 to 16)m ³ /h using air medium	In-house method WI-LAB-09 based on BS EN 14236:2018 and OIML R137-2012 parts 1 and 2
Electronic Metering Device	Accelerated Life Testing (Temperature and Humidity)	In-house method WI-REL-04 & -08 based on IEC 62059-31-1 (Edition 1.0 2008-10)

Signatories:

1. Dr. Li Xiao Qian
2. Mohamed Fairuz Bin Suleiman (*only for)
3. Zakiah Farhanah Multaza (*only for)